

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
 Data File : VU059089.D
 Acq On : 30 May 2024 13:19
 Operator : MD/SY
 Sample : P2589-10ME
 Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH6C4ME

Quant Time: May 31 03:21:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 03:18:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	210551	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	214693	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	98886	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	68234	54.126	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.260%
7) Chloroethane-d5	1.875	69	60914	56.407	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.820%
11) 1,1-Dichloroethene-d2	2.544	63	89958	36.151	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.300%
21) 2-Butanone-d5	4.618	46	115806	97.459	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.460%
24) Chloroform-d	5.052	84	129589	51.637	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.280%
26) 1,2-Dichloroethane-d4	5.692	65	87775	57.064	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.120%
32) Benzene-d6	5.714	84	265671	53.916	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.840%
36) 1,2-Dichloropropane-d6	6.679	67	90002	52.974	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.940%
41) Toluene-d8	7.891	98	225920	54.009	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.020%
43) trans-1,3-Dichloroprop...	8.174	79	38155	55.558	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	111.120%
47) 2-Hexanone-d5	8.634	63	69311	88.219	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.220%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	138837	48.284	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.560%
66) 1,2-Dichlorobenzene-d4	12.190	152	83481	56.570	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
Data File : VU059089.D
Acq On : 30 May 2024 13:19
Operator : MD/SY
Sample : P2589-10ME
Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH6C4ME

Quant Time: May 31 03:21:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 03:18:55 2024
Response via : Initial Calibration

