

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
 Data File : VU059098.D
 Acq On : 30 May 2024 16:44
 Operator : MD/SY
 Sample : P2586-05ME
 Misc : 4.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4D18ME

Quant Time: May 31 03:23:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 03:18:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	204949	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	207785	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	94348	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	60873	49.606	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.220%
7) Chloroethane-d5	1.875	69	49851	47.424	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.840%
11) 1,1-Dichloroethene-d2	2.544	63	81893	33.810	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.620%
21) 2-Butanone-d5	4.618	46	104307	90.181	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.180%
24) Chloroform-d	5.049	84	120855	49.473	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.940%
26) 1,2-Dichloroethane-d4	5.689	65	81373	54.348	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.700%
32) Benzene-d6	5.714	84	242998	50.954	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.900%
36) 1,2-Dichloropropane-d6	6.679	67	81718	49.697	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.400%
41) Toluene-d8	7.891	98	206086	50.905	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.820%
43) trans-1,3-Dichloroprop...	8.174	79	34526	51.945	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.880%
47) 2-Hexanone-d5	8.637	63	65257	85.821	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.820%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	127761	45.909	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.820%
66) 1,2-Dichlorobenzene-d4	12.190	152	74469	52.890	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.780%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
Data File : VU059098.D
Acq On : 30 May 2024 16:44
Operator : MD/SY
Sample : P2586-05ME
Misc : 4.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4D18ME

Quant Time: May 31 03:23:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 03:18:55 2024
Response via : Initial Calibration

