

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
 Data File : VU048739.D
 Acq On : 31 May 2022 14:45
 Operator : SY/MD
 Sample : N3056-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-13D

Quant Time: Jun 01 02:17:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	252374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	429743	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	402307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	197071	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	199953	52.756	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.520%
35) Dibromofluoromethane	5.292	113	149594	48.793	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	7.899	98	531613	48.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	10.632	95	196126	47.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.380%
Target Compounds						
						Qvalue
16) Acetone	2.629	43	3520	1.822	ug/l	95
30) Chloroform	5.092	83	3858	0.669	ug/l	95
64) Tetrachloroethene	8.552	164	3950	0.312	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
Data File : VU048739.D
Acq On : 31 May 2022 14:45
Operator : SY/MD
Sample : N3056-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-13D

Quant Time: Jun 01 02:17:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

