

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
 Data File : VU048744.D
 Acq On : 31 May 2022 17:31
 Operator : SY/MD
 Sample : N3033-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-22

Quant Time: Jun 01 02:19:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	237452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	423878	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	392936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	201027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.704	65	185378	51.984	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.960%
35) Dibromofluoromethane	5.292	113	139767	46.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.440%
50) Toluene-d8	7.900	98	512015	47.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.460%
62) 4-Bromofluorobenzene	10.636	95	190779	46.536	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.080%
Target Compounds						
						Qvalue
16) Acetone	2.626	43	5191	2.856	ug/l	95
30) Chloroform	5.093	83	1908	0.352	ug/l	94
68) m/p-Xylenes	9.700	106	2207	0.403	ug/l	94
84) 1,2,4-Trimethylbenzene	11.475	105	2502	1.388	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
Data File : VU048744.D
Acq On : 31 May 2022 17:31
Operator : SY/MD
Sample : N3033-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-22

Quant Time: Jun 01 02:19:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

