

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
 Data File : VU048756.D
 Acq On : 31 May 2022 22:39
 Operator : SY/MD
 Sample : VU0531MBL02
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0531MBL02

Quant Time: Jun 01 02:24:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	211750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	375105	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	352725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	174344	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	174507	54.875	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.760%
35) Dibromofluoromethane	5.292	113	136337	50.946	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.900%
50) Toluene-d8	7.899	98	454392	47.368	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.740%
62) 4-Bromofluorobenzene	10.632	95	166720	45.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
Data File : VU048756.D
Acq On : 31 May 2022 22:39
Operator : SY/MD
Sample : VU0531MBL02
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0531MBL02

Quant Time: Jun 01 02:24:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

