

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU053122\  
 Data File : VU048764.D  
 Acq On : 01 Jun 2022 02:12  
 Operator : SY/MD  
 Sample : N3099-06  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 40 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 20220429-SPLP-AMENDED-COMPOSITE-I

Quant Time: Jun 01 05:45:39 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U052622W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri May 27 06:42:31 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-----------------------------|--------|----------------|----------|--------|----------|----------|
| -----                       |        |                |          |        |          |          |
| Internal Standards          |        |                |          |        |          |          |
| 1) Pentafluorobenzene       | 5.375  | 168            | 206572   | 50.000 | ug/l     | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.250  | 114            | 364012   | 50.000 | ug/l     | 0.00     |
| 63) Chlorobenzene-d5        | 9.417  | 117            | 345835   | 50.000 | ug/l     | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.812 | 152            | 171397   | 50.000 | ug/l     | 0.00     |
| System Monitoring Compounds |        |                |          |        |          |          |
| 33) 1,2-Dichloroethane-d4   | 5.703  | 65             | 169977   | 54.790 | ug/l     | 0.00     |
| Spiked Amount               | 50.000 | Range 78 - 117 | Recovery | =      | 109.580% |          |
| 35) Dibromofluoromethane    | 0.000  | 113            | 0d       | 0.000  | ug/l     |          |
| Spiked Amount               | 50.000 | Range 75 - 124 | Recovery | =      | 0.000%#  |          |
| 50) Toluene-d8              | 7.899  | 98             | 441033   | 47.377 | ug/l     | 0.00     |
| Spiked Amount               | 50.000 | Range 92 - 112 | Recovery | =      | 94.760%  |          |
| 62) 4-Bromofluorobenzene    | 10.632 | 95             | 160587   | 45.613 | ug/l     | 0.00     |
| Spiked Amount               | 50.000 | Range 83 - 123 | Recovery | =      | 91.220%  |          |
| Target Compounds            |        |                |          |        |          |          |
|                             |        |                |          |        | Qvalue   |          |
| 16) Acetone                 | 2.626  | 43             | 44009    | 27.836 | ug/l     | 99       |
| 20) Methylene Chloride      | 3.047  | 84             | 5098     | 1.896  | ug/l     | 95       |
| 25) 2-Butanone              | 4.713  | 43             | 7030     | 3.095  | ug/l     | 90       |
| -----                       |        |                |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU053122\  
Data File : VU048764.D  
Acq On : 01 Jun 2022 02:12  
Operator : SY/MD  
Sample : N3099-06  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 40 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
20220429-SPLP-AMENDED-COMPOSITE-I

Quant Time: Jun 01 05:45:39 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U052622W.M  
Quant Title : SW846 8260  
QLast Update : Fri May 27 06:42:31 2022  
Response via : Initial Calibration

