

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
 Data File : VU048768.D
 Acq On : 01 Jun 2022 03:46
 Operator : SY/MD
 Sample : N3099-02 20X
 Misc : 3.71g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220420-COMPOSITE-I

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/01/2022
 Supervised By : Mahesh Dadoda 06/02/2022

Quant Time: Jun 01 05:47:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.362	168	298264	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.243	114	504459	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	484842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	241999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	169695	37.884	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	75.760%#		
35) Dibromofluoromethane	5.292	113	127114	35.320	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	70.640%#		
50) Toluene-d8	7.896	98	466847	36.187	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	72.380%#		
62) 4-Bromofluorobenzene	10.639	95	176406	36.156	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	72.320%#		
Target Compounds					Qvalue	
16) Acetone	2.662	43	6061	2.655	ug/l	95
17) Carbon Disulfide	2.719	76	9544	1.077	ug/l	99
18) Methyl Acetate	2.957	43	28791	5.023	ug/l	100
25) 2-Butanone	4.739	43	67290	20.519	ug/l	94
29) Tetrahydrofuran	5.073	42	4957m	2.400	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
Data File : VU048768.D
Acq On : 01 Jun 2022 03:46
Operator : SY/MD
Sample : N3099-02 20X
Misc : 3.71g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220420-COMPOSITE-I

Quant Time: Jun 01 05:47:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/01/2022
Supervised By : Mahesh Dadoda 06/02/2022

