

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
 Data File : VU048771.D
 Acq On : 01 Jun 2022 04:57
 Operator : SY/MD
 Sample : N3075-03 20X
 Misc : 6.11g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS-3

Quant Time: Jun 01 05:48:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/01/2022
 Supervised By :Mahesh Dadoda 06/02/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.363	168	293553	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.244	114	493879	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	461318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	234327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	201722	45.756	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.520%		
35) Dibromofluoromethane	5.295	113	155684	44.185	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.380%		
50) Toluene-d8	7.896	98	641186	50.766	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.540%		
62) 4-Bromofluorobenzene	10.639	95	255135m	53.413	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	106.820%		
Target Compounds						
					Qvalue	
16) Acetone	2.655	43	5658	2.518	ug/l #	88
25) 2-Butanone	4.755	43	6805m	2.108	ug/l	
31) Cyclohexane	5.356	56	825581	148.597	ug/l #	87
39) Methylcyclohexane	6.752	83	3088139	572.729	ug/l	100
40) Benzene	5.761	78	36743	2.420	ug/l	99
52) Toluene	7.970	92	14401	1.586	ug/l	96
67) Ethyl Benzene	9.571	91	969594	58.920	ug/l	98
68) m/p-Xylenes	9.697	106	188336	29.262	ug/l	96
69) o-Xylene	10.102	106	48349	7.646	ug/l	79
73) Isopropylbenzene	10.488	105	749409	49.538	ug/l	100
78) n-propylbenzene	10.909	91	1628805	90.357	ug/l	99
80) 1,3,5-Trimethylbenzene	11.092	105	1337835	101.692	ug/l	99
83) tert-Butylbenzene	11.423	119	39089	3.096	ug/l	95
84) 1,2,4-Trimethylbenzene	11.472	105	1214096	84.785	ug/l	100
85) sec-Butylbenzene	11.648	105	452806	28.666	ug/l	97
86) p-Isopropyltoluene	11.796	119	504451	37.187	ug/l	97
89) n-Butylbenzene	12.208	91	1091055	79.028	ug/l #	37
95) Naphthalene	14.089	128	2868129	150.155	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053122\
Data File : VU048771.D
Acq On : 01 Jun 2022 04:57
Operator : SY/MD
Sample : N3075-03 20X
Misc : 6.11g/10.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS-3

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/01/2022
Supervised By : Mahesh Dadoda 06/02/2022

Quant Time: Jun 01 05:48:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

