

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054334.D
 Acq On : 31 May 2023 13:51
 Operator : JC/MD
 Sample : 02936-21DL 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-02-20230524DL

Quant Time: Jun 01 05:14:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	273233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	448171	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	422188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	198174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	187761	52.421	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.840%		
35) Dibromofluoromethane	5.282	113	153162	53.068	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.140%		
50) Toluene-d8	7.893	98	562254	51.217	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.440%		
62) 4-Bromofluorobenzene	10.629	95	191907	45.085	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.160%		
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.575	101	859	0.303	ug/l #	82
20) Methylene Chloride	3.047	84	1929	0.521	ug/l #	86
27) cis-1,2-Dichloroethene	4.661	96	50008	14.287	ug/l	95
44) Trichloroethene	6.536	130	243551	67.382	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054334.D
Acq On : 31 May 2023 13:51
Operator : JC/MD
Sample : 02936-21DL 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-02-20230524DL

Quant Time: Jun 01 05:14:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

