

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054337.D
 Acq On : 31 May 2023 15:01
 Operator : JC/MD
 Sample : 02936-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PDB-BLANK-20230524

Quant Time: Jun 01 05:21:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	264362	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	436086	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	414662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	197680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	180183	51.993	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.980%	
35) Dibromofluoromethane	5.282	113	147585	52.553	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.100%	
50) Toluene-d8	7.893	98	545764	51.093	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.180%	
62) 4-Bromofluorobenzene	10.629	95	191034	46.123	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.632	43	11738	5.909	ug/l	94
20) Methylene Chloride	3.038	84	924	0.258	ug/l #	70
27) cis-1,2-Dichloroethene	4.668	96	1848	0.546	ug/l	94
37) Ethyl Acetate	4.812	43	9932	1.848	ug/l #	91
44) Trichloroethene	6.536	130	1238	0.352	ug/l	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054337.D
Acq On : 31 May 2023 15:01
Operator : JC/MD
Sample : 02936-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PDB-BLANK-20230524

Quant Time: Jun 01 05:21:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

