

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054338.D
 Acq On : 31 May 2023 15:25
 Operator : JC/MD
 Sample : 02936-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-01-20230524

Quant Time: Jun 01 05:21:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	264056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	446673	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	418824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	195425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	182070	52.598	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.200%
35) Dibromofluoromethane	5.282	113	148771	51.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.440%
50) Toluene-d8	7.893	98	546914	49.987	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.980%
62) 4-Bromofluorobenzene	10.626	95	187778	44.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.520%
Target Compounds						
						Qvalue
16) Acetone	2.629	43	7918	3.991	ug/l #	86
20) Methylene Chloride	3.041	84	1134	0.317	ug/l #	91
30) Chloroform	5.083	83	66073	11.629	ug/l	100
44) Trichloroethene	6.536	130	954	0.265	ug/l	78
47) Bromodichloromethane	7.099	83	8254	1.784	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054338.D
Acq On : 31 May 2023 15:25
Operator : JC/MD
Sample : 02936-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-01-20230524

Quant Time: Jun 01 05:21:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

