

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054339.D
 Acq On : 31 May 2023 15:49
 Operator : JC/MD
 Sample : 02936-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 91BR-20230523

Quant Time: Jun 01 05:22:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	268287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	454658	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	417202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	193146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	182099	51.777	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.560%	
35) Dibromofluoromethane	5.282	113	151100	51.607	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.220%	
50) Toluene-d8	7.893	98	554523	49.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.580%	
62) 4-Bromofluorobenzene	10.629	95	187952	43.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.060%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.382	85	929	0.310	ug/l	98
16) Acetone	2.629	43	1767	0.877	ug/l	98
20) Methylene Chloride	3.041	84	1391	0.382	ug/l	95
21) trans-1,2-Dichloroethene	3.346	96	1520	0.492	ug/l	97
27) cis-1,2-Dichloroethene	4.665	96	12794	3.722	ug/l	95
44) Trichloroethene	6.539	130	12290	3.352	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054339.D
Acq On : 31 May 2023 15:49
Operator : JC/MD
Sample : 02936-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
91BR-20230523

Quant Time: Jun 01 05:22:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

