

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054340.D
 Acq On : 31 May 2023 16:12
 Operator : JC/MD
 Sample : 02912-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB02-20230523

Quant Time: Jun 01 05:22:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	268349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	455644	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	426622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	203202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	187024	53.165	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 106.340%		
35) Dibromofluoromethane	5.282	113	152854	52.093	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.180%		
50) Toluene-d8	7.893	98	563512	50.490	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.980%		
62) 4-Bromofluorobenzene	10.629	95	192932	44.582	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 89.160%		
Target Compounds						
					Qvalue	
16) Acetone	2.636	43	4016	1.992	ug/l	91
20) Methylene Chloride	3.038	84	1385	0.381	ug/l #	74
30) Chloroform	5.086	83	5909	1.023	ug/l	90
44) Trichloroethene	6.539	130	795	0.216	ug/l	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054340.D
Acq On : 31 May 2023 16:12
Operator : JC/MD
Sample : 02912-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB02-20230523

Quant Time: Jun 01 05:22:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

