

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054345.D
 Acq On : 31 May 2023 18:11
 Operator : JC/MD
 Sample : PB153123ZHE#04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#04

Quant Time: Jun 01 05:24:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	241648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	420448	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	398957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	186118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	169370	53.467	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 106.940%		
35) Dibromofluoromethane	5.282	113	140983	52.070	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.140%		
50) Toluene-d8	7.893	98	520648	50.554	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.100%		
62) 4-Bromofluorobenzene	10.626	95	177357	44.414	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 88.820%		
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	23174	12.763	ug/l	99
20) Methylene Chloride	3.041	84	20864	6.369	ug/l	99
37) Ethyl Acetate	4.800	43	43714	8.434	ug/l #	92
43) Isopropyl Acetate	5.903	43	166512	23.292	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054345.D
Acq On : 31 May 2023 18:11
Operator : JC/MD
Sample : PB153123ZHE#04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#04

Quant Time: Jun 01 05:24:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

