

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054348.D
 Acq On : 31 May 2023 19:23
 Operator : JC/MD
 Sample : PB153123ZHE#07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#07

Quant Time: Jun 01 05:25:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	247331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	436639	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	404180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	189492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	176114	54.318	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.640%
35) Dibromofluoromethane	5.282	113	142498	50.678	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.360%
50) Toluene-d8	7.893	98	531714	49.714	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.420%
62) 4-Bromofluorobenzene	10.626	95	179617	43.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.620%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	26078	14.032	ug/l	100
20) Methylene Chloride	3.041	84	19881	5.930	ug/l	90
37) Ethyl Acetate	4.800	43	46105	8.566	ug/l #	92
43) Isopropyl Acetate	5.906	43	175590	23.651	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054348.D
Acq On : 31 May 2023 19:23
Operator : JC/MD
Sample : PB153123ZHE#07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#07

Quant Time: Jun 01 05:25:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

