

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054349.D
 Acq On : 31 May 2023 19:46
 Operator : JC/MD
 Sample : PB153123ZHE#08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#08

Quant Time: Jun 01 05:26:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	242773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	436373	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	404815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	188875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	171877	54.007	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	108.020%	
35) Dibromofluoromethane	5.282	113	141811	50.464	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.920%	
50) Toluene-d8	7.893	98	528447	49.439	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.880%	
62) 4-Bromofluorobenzene	10.629	95	179160	43.228	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	86.460%	
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	24023	13.169	ug/l	97
20) Methylene Chloride	3.041	84	20626	6.267	ug/l	95
37) Ethyl Acetate	4.800	43	43796	8.142	ug/l #	92
43) Isopropyl Acetate	5.906	43	168350	22.689	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054349.D
Acq On : 31 May 2023 19:46
Operator : JC/MD
Sample : PB153123ZHE#08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#08

Quant Time: Jun 01 05:26:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

