

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054350.D
 Acq On : 31 May 2023 20:10
 Operator : JC/MD
 Sample : PB153123ZHE#09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#09

Quant Time: Jun 01 05:26:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	251949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	435218	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	406521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	178516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	172821	52.326	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.660%		
35) Dibromofluoromethane	5.282	113	145190	51.804	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.600%		
50) Toluene-d8	7.893	98	535354	50.218	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.440%		
62) 4-Bromofluorobenzene	10.629	95	175618	42.486	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 84.980%		
Target Compounds						
16) Acetone	2.626	43	25383	13.408	ug/l	Qvalue 96
20) Methylene Chloride	3.038	84	21557	6.312	ug/l	97
37) Ethyl Acetate	4.800	43	44925	8.374	ug/l #	92
43) Isopropyl Acetate	5.903	43	174177	23.537	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054350.D
Acq On : 31 May 2023 20:10
Operator : JC/MD
Sample : PB153123ZHE#09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#09

Quant Time: Jun 01 05:26:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

