

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054351.D
 Acq On : 31 May 2023 20:34
 Operator : JC/MD
 Sample : PB153123ZHE#10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#10

Quant Time: Jun 01 05:26:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	253292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	449413	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	421316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	194729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	177468	53.448	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.900%
35) Dibromofluoromethane	5.282	113	145488	50.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.540%
50) Toluene-d8	7.893	98	555461	50.458	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.920%
62) 4-Bromofluorobenzene	10.629	95	186654	43.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.460%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	28529	14.989	ug/l	96
18) Methyl Acetate	2.951	43	7865	1.132	ug/l	93
20) Methylene Chloride	3.041	84	21329	6.212	ug/l	93
37) Ethyl Acetate	4.800	43	48307	8.720	ug/l #	92
43) Isopropyl Acetate	5.906	43	180210	23.583	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054351.D
Acq On : 31 May 2023 20:34
Operator : JC/MD
Sample : PB153123ZHE#10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#10

Quant Time: Jun 01 05:26:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

