

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054361.D
 Acq On : 01 Jun 2023 00:54
 Operator : JC/MD
 Sample : PB153123ZHE#12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB153123ZHE#12

Quant Time: Jun 01 06:40:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	257262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	448468	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	426267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	192120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	182142	54.009	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.020%		
35) Dibromofluoromethane	5.282	113	150535	52.124	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.240%		
50) Toluene-d8	7.893	98	561151	51.083	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.160%		
62) 4-Bromofluorobenzene	10.626	95	187179	43.945	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 87.880%		
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	27345	14.146	ug/l	98
20) Methylene Chloride	3.041	84	23362	6.699	ug/l #	91
37) Ethyl Acetate	4.803	43	47742	8.636	ug/l #	92
43) Isopropyl Acetate	5.906	43	178131	23.360	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054361.D
Acq On : 01 Jun 2023 00:54
Operator : JC/MD
Sample : PB153123ZHE#12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB153123ZHE#12

Quant Time: Jun 01 06:40:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

