

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054362.D
 Acq On : 01 Jun 2023 01:18
 Operator : JC/MD
 Sample : 02972-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-11-0A

Quant Time: Jun 01 06:40:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	258999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	448790	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	429597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	198669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	184224	54.260	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.520%	
35) Dibromofluoromethane	5.282	113	149873	51.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.720%	
50) Toluene-d8	7.893	98	561373	51.066	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.140%	
62) 4-Bromofluorobenzene	10.626	95	189330	44.418	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.840%	
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	37694	19.368	ug/l	99
20) Methylene Chloride	3.044	84	22364	6.370	ug/l	95
37) Ethyl Acetate	4.803	43	42657	7.711	ug/l #	89
40) Benzene	5.771	78	22877	1.767	ug/l	100
43) Isopropyl Acetate	5.906	43	175323	22.975	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054362.D
Acq On : 01 Jun 2023 01:18
Operator : JC/MD
Sample : 02972-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-11-0A

Quant Time: Jun 01 06:40:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

