

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054363.D
 Acq On : 01 Jun 2023 01:42
 Operator : JC/MD
 Sample : 02972-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-12-0A

Quant Time: Jun 01 06:41:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	256684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	449879	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	428504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	206141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	182238	54.159	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.320%	
35) Dibromofluoromethane	5.282	113	151013	52.125	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.260%	
50) Toluene-d8	7.893	98	560799	50.891	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.780%	
62) 4-Bromofluorobenzene	10.626	95	192078	44.953	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.900%	
Target Compounds						
						Qvalue
16) Acetone	2.626	43	55851	28.957	ug/l	94
20) Methylene Chloride	3.041	84	23060	6.627	ug/l	98
37) Ethyl Acetate	4.803	43	46759	8.432	ug/l #	94
43) Isopropyl Acetate	5.903	43	165752	21.669	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054363.D
Acq On : 01 Jun 2023 01:42
Operator : JC/MD
Sample : 02972-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-12-0A

Quant Time: Jun 01 06:41:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

