

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054367.D
 Acq On : 01 Jun 2023 03:16
 Operator : JC/MD
 Sample : 02979-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RBR044-VNJ224

Quant Time: Jun 01 06:42:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	260830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	461067	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	436819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	206351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	179672	52.548	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.100%
35) Dibromofluoromethane	5.282	113	148439	49.993	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	7.893	98	570478	50.513	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.020%
62) 4-Bromofluorobenzene	10.626	95	197258	45.045	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.100%
Target Compounds						
						Qvalue
16) Acetone	2.629	43	42605	21.738	ug/l	97
20) Methylene Chloride	3.041	84	8963	2.535	ug/l	89
37) Ethyl Acetate	4.803	43	51107	8.992	ug/l #	92
43) Isopropyl Acetate	5.903	43	197416	25.182	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054367.D
Acq On : 01 Jun 2023 03:16
Operator : JC/MD
Sample : 02979-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RBR044-VNJ224

Quant Time: Jun 01 06:42:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

