

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
 Data File : VU054369.D
 Acq On : 01 Jun 2023 04:04
 Operator : JC/MD
 Sample : 02981-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-4

Quant Time: Jun 01 06:43:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	260272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	455910	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	430530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	199541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	183947	53.913	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.820%		
35) Dibromofluoromethane	5.282	113	150634	51.307	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.620%		
50) Toluene-d8	7.893	98	574335	51.429	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.860%		
62) 4-Bromofluorobenzene	10.626	95	194277	44.867	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 89.740%		
Target Compounds						
					Qvalue	
16) Acetone	2.630	43	27569	14.097	ug/l	94
20) Methylene Chloride	3.041	84	21206	6.010	ug/l	91
37) Ethyl Acetate	4.803	43	46679	8.306	ug/l #	92
43) Isopropyl Acetate	5.903	43	175160	22.596	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053123\
Data File : VU054369.D
Acq On : 01 Jun 2023 04:04
Operator : JC/MD
Sample : 02981-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-4

Quant Time: Jun 01 06:43:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

