

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
 Data File : VU048783.D
 Acq On : 01 Jun 2022 15:42
 Operator : SY/MD
 Sample : N3075-02 50X
 Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS-2

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/02/2022
 Supervised By :Semsettin Yesilyurt 06/02/2022

Quant Time: Jun 02 05:04:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	269459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	454145	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	435014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	226455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	204504	50.535	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.080%			
35) Dibromofluoromethane	5.292	113	163895	50.585	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.180%			
50) Toluene-d8	7.899	98	589695	50.774	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.540%			
62) 4-Bromofluorobenzene	10.632	95	223629	50.913	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.820%			
Target Compounds					Qvalue	
31) Cyclohexane	5.385	56	27102	5.314	ug/l #	85
39) Methylcyclohexane	6.764	83	49636	10.011	ug/l	98
40) Benzene	5.777	78	30505	2.185	ug/l	96
52) Toluene	7.970	92	3046	0.365	ug/l	87
67) Ethyl Benzene	9.571	91	66591	4.291	ug/l	98
68) m/p-Xylenes	9.697	106	31275	5.153	ug/l	98
69) o-Xylene	10.102	106	2680	0.449	ug/l	99
73) Isopropylbenzene	10.484	105	12604	0.862	ug/l	98
78) n-propylbenzene	10.906	91	27202	1.561	ug/l	96
80) 1,3,5-Trimethylbenzene	11.089	105	40723	3.203	ug/l	98
84) 1,2,4-Trimethylbenzene	11.468	105	119037	9.668	ug/l	98
85) sec-Butylbenzene	11.645	105	5657	0.371	ug/l	97
86) p-Isopropyltoluene	11.793	119	7570m	0.577	ug/l	
89) n-Butylbenzene	12.201	91	14755	3.261	ug/l #	30
95) Naphthalene	14.089	128	55066	5.465	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
Data File : VU048783.D
Acq On : 01 Jun 2022 15:42
Operator : SY/MD
Sample : N3075-02 50X
Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS-2

Quant Time: Jun 02 05:04:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/02/2022
Supervised By :Semsettin Yesilyurt 06/02/2022

