

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
 Data File : VU048788.D
 Acq On : 01 Jun 2022 18:30
 Operator : SY/MD
 Sample : N3099-05RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0602

Quant Time: Jun 02 05:07:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	279967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	473954	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	440847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	227520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	202683	48.205	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.420%
35) Dibromofluoromethane	5.292	113	690	0.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.400%#
50) Toluene-d8	7.899	98	563688	46.506	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.020%
62) 4-Bromofluorobenzene	10.632	95	215073	46.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.840%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.604	62	8495	2.223	ug/l	96
16) Acetone	2.623	43	73824	34.453	ug/l	98
20) Methylene Chloride	3.047	84	6921	1.899	ug/l	89
24) 1,1-Dichloroethane	3.874	63	26884	4.205	ug/l	98
25) 2-Butanone	4.713	43	14131	4.591	ug/l	98
27) cis-1,2-Dichloroethene	4.665	96	21463	5.689	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
Data File : VU048788.D
Acq On : 01 Jun 2022 18:30
Operator : SY/MD
Sample : N3099-05RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0602

Quant Time: Jun 02 05:07:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

