

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
 Data File : VU048789.D
 Acq On : 01 Jun 2022 18:53
 Operator : SY/MD
 Sample : N3099-06RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0603

Quant Time: Jun 02 05:07:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	274565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	467365	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	440589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	221372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	213207	51.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.420%
35) Dibromofluoromethane	5.289	113	262	0.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.160%#
50) Toluene-d8	7.899	98	596350	49.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	10.632	95	226074	50.014	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	56063	26.679	ug/l	97
20) Methylene Chloride	3.044	84	6492	1.817	ug/l	93
25) 2-Butanone	4.710	43	10019	3.319	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
Data File : VU048789.D
Acq On : 01 Jun 2022 18:53
Operator : SY/MD
Sample : N3099-06RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0603

Quant Time: Jun 02 05:07:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

