

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
 Data File : VU048793.D
 Acq On : 01 Jun 2022 20:27
 Operator : SY/MD
 Sample : N3075-01ME 20X
 Misc : 6.96g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS-1ME

Quant Time: Jun 02 08:36:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	282712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	476643	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	439540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	235528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	201553	47.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.940%
35) Dibromofluoromethane	5.288	113	168726	49.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.240%
50) Toluene-d8	7.899	98	628054	51.524	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.040%
62) 4-Bromofluorobenzene	10.632	95	241407	52.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.740%
Target Compounds						
						Qvalue
31) Cyclohexane	5.388	56	345320	64.538	ug/l	98
39) Methylcyclohexane	6.764	83	694946	133.546	ug/l	99
40) Benzene	5.774	78	392571	26.793	ug/l	100
52) Toluene	7.970	92	1035018	118.098	ug/l	100
67) Ethyl Benzene	9.571	91	868211	55.373	ug/l	98
68) m/p-Xylenes	9.690	106	1133163	184.786	ug/l	98
69) o-Xylene	10.099	106	439021	72.870	ug/l	99
73) Isopropylbenzene	10.484	105	138359	9.099	ug/l	99
78) n-propylbenzene	10.906	91	299637	16.537	ug/l	99
80) 1,3,5-Trimethylbenzene	11.086	105	512217	38.736	ug/l	100
84) 1,2,4-Trimethylbenzene	11.465	105	1437590	99.669	ug/l	99
85) sec-Butylbenzene	11.642	105	53588	3.375	ug/l	100
86) p-Isopropyltoluene	11.793	119	72040	5.284	ug/l	96
89) n-Butylbenzene	12.195	91	149701	12.675	ug/l #	22
95) Naphthalene	14.086	128	476240	26.920	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060122\
 Data File : VU048793.D
 Acq On : 01 Jun 2022 20:27
 Operator : SY/MD
 Sample : N3075-01ME 20X
 Misc : 6.96g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS-1ME

Quant Time: Jun 02 08:36:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

