

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048807.D
 Acq On : 02 Jun 2022 16:35
 Operator : SY/MD
 Sample : N3143-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220425-SPLP-AMENDED-COMP-E

Quant Time: Jun 03 01:16:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	348838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	569098	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	529282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	268840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	231274	44.146	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.300%
35) Dibromofluoromethane	5.289	113	88803	21.872	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	43.740%#
50) Toluene-d8	7.896	98	694987	47.753	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.500%
62) 4-Bromofluorobenzene	10.632	95	253245	46.010	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.020%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	81351	30.470	ug/l	98
20) Methylene Chloride	3.041	84	5871	1.293	ug/l	97
25) 2-Butanone	4.707	43	21444	5.591	ug/l	92
84) 1,2,4-Trimethylbenzene	11.468	105	19071	2.332	ug/l	98
95) Naphthalene	14.086	128	618914	30.299	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048807.D
Acq On : 02 Jun 2022 16:35
Operator : SY/MD
Sample : N3143-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220425-SPLP-AMENDED-COMP-E

Quant Time: Jun 03 01:16:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

