

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048809.D
 Acq On : 02 Jun 2022 17:22
 Operator : SY/MD
 Sample : N3112-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB07-20220527

Quant Time: Jun 03 01:17:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	358392	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	587925	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	535592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	249604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	242235	45.005	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.020%
35) Dibromofluoromethane	5.289	113	188114	44.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.700%
50) Toluene-d8	7.896	98	723765	48.137	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.280%
62) 4-Bromofluorobenzene	10.632	95	235359	41.391	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.780%#
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	3877	1.413	ug/l	92
18) Methyl Acetate	2.951	43	2679	0.389	ug/l	98
95) Naphthalene	14.089	128	67326	5.786	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048809.D
Acq On : 02 Jun 2022 17:22
Operator : SY/MD
Sample : N3112-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB07-20220527

Quant Time: Jun 03 01:17:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

