

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048810.D
 Acq On : 02 Jun 2022 17:45
 Operator : SY/MD
 Sample : N3112-23
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB01-20220527

Quant Time: Jun 03 01:17:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	353969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	583056	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	536073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	234009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	239055	44.970	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.940%
35) Dibromofluoromethane	5.289	113	186155	44.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.500%
50) Toluene-d8	7.896	98	701476	47.045	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.080%
62) 4-Bromofluorobenzene	10.632	95	247079	43.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.620%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	32061	11.834	ug/l	99
20) Methylene Chloride	3.044	84	2792	0.606	ug/l #	88
25) 2-Butanone	4.713	43	7304	1.877	ug/l	93
95) Naphthalene	14.092	128	41095	4.651	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048810.D
Acq On : 02 Jun 2022 17:45
Operator : SY/MD
Sample : N3112-23
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB01-20220527

Quant Time: Jun 03 01:17:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

