

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048813.D
 Acq On : 02 Jun 2022 18:56
 Operator : SY/MD
 Sample : N3112-13 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 INTERCEPTOR-SUMP-20220527

Quant Time: Jun 03 01:18:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	346554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	564392	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	521037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	234629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	234452	45.047	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.100%
35) Dibromofluoromethane	5.289	113	181458	45.066	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.140%
50) Toluene-d8	7.896	98	692132	47.953	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.900%
62) 4-Bromofluorobenzene	10.632	95	234313	42.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.860%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.578	96	2499	0.708	ug/l #	70
21) trans-1,2-Dichloroethene	3.353	96	3755	0.931	ug/l	86
27) cis-1,2-Dichloroethene	4.665	96	109325	23.409	ug/l	87
44) Trichloroethene	6.542	130	451739	101.065	ug/l	98
95) Naphthalene	14.092	128	15554	3.332	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048813.D
Acq On : 02 Jun 2022 18:56
Operator : SY/MD
Sample : N3112-13 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
INTERCEPTOR-SUMP-20220527

Quant Time: Jun 03 01:18:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

