

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048818.D
 Acq On : 02 Jun 2022 20:55
 Operator : SY/MD
 Sample : N3112-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SPRING-20220527

Quant Time: Jun 03 01:20:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	352395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	573294	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	505306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	221465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	240606	45.463	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.920%		
35) Dibromofluoromethane	5.288	113	182974	44.737	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.480%		
50) Toluene-d8	7.899	98	700736	47.795	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.600%		
62) 4-Bromofluorobenzene	10.632	95	232748	41.976	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.960%		
Target Compounds						
16) Acetone	2.639	43	4521	1.676	ug/l	98
95) Naphthalene	14.095	128	6886	2.908	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048818.D
Acq On : 02 Jun 2022 20:55
Operator : SY/MD
Sample : N3112-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SPRING-20220527

Quant Time: Jun 03 01:20:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

