

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048829.D
 Acq On : 03 Jun 2022 02:04
 Operator : SY/MD
 Sample : N3112-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GR-OF-20220527

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022
 Supervised By :Semsettin Yesilyurt 06/03/2022

Quant Time: Jun 03 04:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	324092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	540072	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	494770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	250263	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	216432	44.467	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	88.940%		
35) Dibromofluoromethane	5.288	113	167502	43.473	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	86.940%		
50) Toluene-d8	7.896	98	629188	45.555	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	91.120%#		
62) 4-Bromofluorobenzene	10.632	95	231790	44.375	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.760%		
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	8791	3.544	ug/l	92
25) 2-Butanone	4.716	43	4264	1.197	ug/l	97
27) cis-1,2-Dichloroethene	4.671	96	2835m	0.649	ug/l	
44) Trichloroethene	6.539	130	3419	0.799	ug/l	68
95) Naphthalene	14.089	128	55196	5.193	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048829.D
Acq On : 03 Jun 2022 02:04
Operator : SY/MD
Sample : N3112-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GR-OF-20220527

Quant Time: Jun 03 04:05:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022
Supervised By :Semsettin Yesilyurt 06/03/2022

