

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048832.D
 Acq On : 03 Jun 2022 03:15
 Operator : SY/MD
 Sample : N3112-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-118.5N-20220527

Quant Time: Jun 03 04:06:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	334247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	539253	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	491826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	227169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	233294	46.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.960%
35) Dibromofluoromethane	5.288	113	176455	45.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.740%
50) Toluene-d8	7.896	98	663564	48.117	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	10.632	95	229270	43.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.920%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	4314	1.686	ug/l	92
44) Trichloroethene	6.542	130	1746	0.409	ug/l	88
95) Naphthalene	14.095	128	11576	3.147	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048832.D
Acq On : 03 Jun 2022 03:15
Operator : SY/MD
Sample : N3112-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-118.5N-20220527

Quant Time: Jun 03 04:06:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

