

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048834.D
 Acq On : 03 Jun 2022 04:02
 Operator : SY/MD
 Sample : N3112-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-121.5N-20220527

Quant Time: Jun 03 04:35:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	329265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	523613	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	495786	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	224559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	233386	47.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.400%
35) Dibromofluoromethane	5.288	113	177880	47.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	7.896	98	665710	49.714	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.420%
62) 4-Bromofluorobenzene	10.632	95	222778	43.990	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.980%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	4455	1.768	ug/l	93
27) cis-1,2-Dichloroethene	4.674	96	2169	0.489	ug/l	86
44) Trichloroethene	6.539	130	2264	0.546	ug/l	79
95) Naphthalene	14.092	128	6912	2.904	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048834.D
Acq On : 03 Jun 2022 04:02
Operator : SY/MD
Sample : N3112-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-121.5N-20220527

Quant Time: Jun 03 04:35:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

