

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048835.D
 Acq On : 03 Jun 2022 04:25
 Operator : SY/MD
 Sample : N3112-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-121.5T-20220527

Quant Time: Jun 03 04:39:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	324735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	531514	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	476793	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	214712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	230274	47.217	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.440%
35) Dibromofluoromethane	5.288	113	176365	46.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.020%
50) Toluene-d8	7.896	98	637777	46.920	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.840%
62) 4-Bromofluorobenzene	10.632	95	215393	41.900	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.800%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	4795	1.929	ug/l	96
27) cis-1,2-Dichloroethene	4.668	96	44651	10.203	ug/l	89
44) Trichloroethene	6.542	130	31051	7.377	ug/l	97
95) Naphthalene	14.095	128	5827	2.860	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048835.D
Acq On : 03 Jun 2022 04:25
Operator : SY/MD
Sample : N3112-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-121.5T-20220527

Quant Time: Jun 03 04:39:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

