

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
 Data File : VU048836.D
 Acq On : 03 Jun 2022 04:49
 Operator : SY/MD
 Sample : N3112-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-125.9T-20220527

Quant Time: Jun 03 06:56:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	324614	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	533665	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	487135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	219191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	229946	47.168	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.340%
35) Dibromofluoromethane	5.289	113	176440	46.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	7.899	98	643415	47.144	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.280%
62) 4-Bromofluorobenzene	10.632	95	221589	42.931	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.860%
Target Compounds						
					Qvalue	
3) Chloromethane	1.520	50	1240	0.259	ug/l	96
16) Acetone	2.626	43	5301	2.134	ug/l	96
95) Naphthalene	14.098	128	4589	2.785	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060222\
Data File : VU048836.D
Acq On : 03 Jun 2022 04:49
Operator : SY/MD
Sample : N3112-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-125.9T-20220527

Quant Time: Jun 03 06:56:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

