

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048854.D
 Acq On : 03 Jun 2022 13:55
 Operator : SY/MD
 Sample : N3103-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 04 04:36:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	301187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	492278	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	438277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	184124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	214881	47.506	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.288	113	162156	46.172	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.340%
50) Toluene-d8	7.896	98	593503	47.143	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.280%
62) 4-Bromofluorobenzene	10.632	95	193016	40.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048854.D
Acq On : 03 Jun 2022 13:55
Operator : SY/MD
Sample : N3103-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 04 04:36:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

