

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048856.D
 Acq On : 03 Jun 2022 14:42
 Operator : SY/MD
 Sample : N3103-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 04 04:37:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	299587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	491917	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	447743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	196612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	214565	47.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.380%
35) Dibromofluoromethane	5.288	113	162716	46.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.740%
50) Toluene-d8	7.896	98	599737	47.673	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.340%
62) 4-Bromofluorobenzene	10.632	95	201385	42.328	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.660%
Target Compounds						
16) Acetone	2.626	43	3417	1.490	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048856.D
Acq On : 03 Jun 2022 14:42
Operator : SY/MD
Sample : N3103-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 04 04:37:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

