

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048857.D
 Acq On : 03 Jun 2022 15:06
 Operator : SY/MD
 Sample : N3103-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 04 04:37:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	306354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	497006	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	446368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	192376	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	222446	48.349	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.700%
35) Dibromofluoromethane	5.292	113	163884	46.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.440%
50) Toluene-d8	7.896	98	596720	46.948	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.900%
62) 4-Bromofluorobenzene	10.632	95	197444	41.075	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.160%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048857.D
Acq On : 03 Jun 2022 15:06
Operator : SY/MD
Sample : N3103-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 04 04:37:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

