

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048858.D
 Acq On : 03 Jun 2022 15:29
 Operator : SY/MD
 Sample : N3112-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-118.5T-20220527

Quant Time: Jun 04 04:37:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	294799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	487764	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	446092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	197051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	224265	50.655	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.300%
35) Dibromofluoromethane	5.289	113	169882	48.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.640%
50) Toluene-d8	7.899	98	613934	49.218	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	10.632	95	204554	43.361	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.720%
Target Compounds						
16) Acetone	2.626	43	3338	1.479	ug/l	92
27) cis-1,2-Dichloroethene	4.668	96	17483	4.401	ug/l	89
44) Trichloroethene	6.542	130	11274	2.919	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048858.D
Acq On : 03 Jun 2022 15:29
Operator : SY/MD
Sample : N3112-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-118.5T-20220527

Quant Time: Jun 04 04:37:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

