

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048863.D
 Acq On : 03 Jun 2022 17:27
 Operator : SY/MD
 Sample : N3112-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP01-20220527

Quant Time: Jun 04 04:39:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	286934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	469122	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	423999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	186178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	217673	50.514	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.020%
35) Dibromofluoromethane	5.288	113	162344	48.507	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	7.899	98	581595	48.478	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.960%
62) 4-Bromofluorobenzene	10.632	95	191837	42.281	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.560%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	3595	1.637	ug/l	93
44) Trichloroethene	6.542	130	1638	0.441	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048863.D
Acq On : 03 Jun 2022 17:27
Operator : SY/MD
Sample : N3112-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP01-20220527

Quant Time: Jun 04 04:39:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

