

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048864.D
 Acq On : 03 Jun 2022 17:51
 Operator : SY/MD
 Sample : N3112-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP02-20220527

Quant Time: Jun 04 04:39:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	278142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	460172	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	422566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	181312	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	215213	51.521	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.040%			
35) Dibromofluoromethane	5.288	113	159607	48.617	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.240%			
50) Toluene-d8	7.899	98	577852	49.103	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.200%			
62) 4-Bromofluorobenzene	10.632	95	191056	42.928	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 85.860%			
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	4869	2.287	ug/l #	80
27) cis-1,2-Dichloroethene	4.671	96	3953	1.055	ug/l	84
44) Trichloroethene	6.542	130	8299	2.277	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048864.D
Acq On : 03 Jun 2022 17:51
Operator : SY/MD
Sample : N3112-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP02-20220527

Quant Time: Jun 04 04:39:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

