

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048866.D
 Acq On : 03 Jun 2022 18:38
 Operator : SY/MD
 Sample : N3170-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220424-SPLP-AMENDED-COMPOSITE-D

Quant Time: Jun 04 04:40:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	274288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	445217	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	407015	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	181077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	202457	49.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.300%
35) Dibromofluoromethane	5.288	113	69375	21.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	43.680%#
50) Toluene-d8	7.896	98	540803	47.498	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.000%
62) 4-Bromofluorobenzene	10.632	95	179800	41.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.520%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	76189	36.293	ug/l	97
20) Methylene Chloride	3.044	84	4794	1.343	ug/l #	88
25) 2-Butanone	4.710	43	14001	4.643	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048866.D
Acq On : 03 Jun 2022 18:38
Operator : SY/MD
Sample : N3170-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220424-SPLP-AMENDED-COMPOSITE-D

Quant Time: Jun 04 04:40:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

