

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048867.D
 Acq On : 03 Jun 2022 19:02
 Operator : SY/MD
 Sample : N3165-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 47BR-20220531

Quant Time: Jun 04 04:40:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	275715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	450553	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	411951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	181970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	222414	53.714	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.420%	
35) Dibromofluoromethane	5.288	113	167539	52.122	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.240%	
50) Toluene-d8	7.896	98	591314	51.319	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.640%	
62) 4-Bromofluorobenzene	10.632	95	192453	44.165	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.320%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.604	62	12191	3.239	ug/l #	57
12) 1,1-Dichloroethene	2.575	96	1037	0.370	ug/l	91
16) Acetone	2.626	43	6850	3.246	ug/l	94
27) cis-1,2-Dichloroethene	4.665	96	123872	33.339	ug/l	88
37) Ethyl Acetate	4.806	43	76598	12.916	ug/l	98
44) Trichloroethene	6.542	130	27913	7.823	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
Data File : VU048867.D
Acq On : 03 Jun 2022 19:02
Operator : SY/MD
Sample : N3165-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
47BR-20220531

Quant Time: Jun 04 04:40:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

