


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060322\
 Data File : VU048847.D
 Acq On : 03 Jun 2022 10:05
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Title : SW846 8260
 Last Update : Fri May 27 06:42:31 2022



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2877

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	11120	PASS
75	95	30	60	51.3	29696	PASS
95	95	100	100	100.0	57912	PASS
96	95	5	9	6.2	3606	PASS
173	174	0.00	2	0.5	203	PASS
174	95	50	100	77.9	45107	PASS
175	174	5	9	7.5	3402	PASS
176	174	95	101	97.7	44051	PASS
177	176	5	9	6.5	2867	PASS