

Data File : VU032445.D
Acq On : 04 Jun 2019 22:27
Operator : JC/SP
Sample : VIBLK40
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jun 05 00:19:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 04 22:54:09 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	108296	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	99342	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	50529	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	30172	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.67	69	25374	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.40%
11) 1,1-Dichloroethene-d2	2.27	63	45569	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	4.18	46	79137	40.27	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.54%
24) Chloroform-d	4.64	84	52967	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
26) 1,2-Dichloroethane-d4	5.30	65	29723	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
32) Benzene-d6	5.33	84	111287	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
36) 1,2-Dichloropropane-d6	6.32	67	34656	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.56	98	102331	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	7.85	79	12661	3.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.80%
46) 2-Hexanone-d5	8.31	63	66685	40.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	28435	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	40844	4.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.80%


Target Compounds					Qvalue
34) Trichloroethene	6.18	95	1267	0.150 ug/L	89
35) Methylcyclohexane	6.32	83	7779	0.574 ug/L #	20
37) 1,2-Dichloropropane	6.32	63	3887	0.513 ug/L #	89
59) 1,2,3-Trichloropropane	11.48	75	5894	1.184 ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU032445.D
Acq On : 04 Jun 2019 22:27
Operator : JC/SP
Sample : VIBLK40
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jun 05 00:19:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 04 22:54:09 2019
Response via : Initial Calibration





#34

Trichloroethene

Concen: 0.150 ug/L

RT: 6.18 min Scan# 1583

Delta R.T. -0.00 min

Lab File: VU032445.D

Acq: 04 Jun 2019 22:27

Instrument :

MSVOA_U

ClientSampleId :

VIBLK40

Tgt Ion: 95 Resp: 1267



Ion Ratio Lower Upper

95 100

97 57.0 44.1 81.9

132 101.6 68.8 127.8

130 81.5 71.0 131.8





Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 96.95 (96.65 to 97.65): VU0

Ion 131.95 (131.65 to 132.65): V

Ion 129.95 (129.65 to 130.65): V



#35

Methylcyclohexane

Concen: 0.574 ug/L

RT: 6.32 min Scan# 1628

Delta R.T. -0.08 min

Lab File: VU032445.D

Acq: 04 Jun 2019 22:27



Tgt Ion: 83 Resp: 7779

Ion Ratio Lower Upper

83 100

55 0.0 60.0 90.0#


98 2.7 37.8 56.8#




Abundance Ion 83.15 (82.85 to 83.85): VU0

Ion 55.10 (54.80 to 55.80): VU0

Ion 98.15 (97.85 to 98.85): VU0









#37
 1,2-Dichloropropane
 Concen: 0.513 ug/L
 RT: 6.32 min Scan# 1628
 Delta R.T. -0.10 min
 Lab File: VU032445.D
 Acq: 04 Jun 2019 22:27

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Tgt Ion: 63 Resp: 3887
 Ion Ratio Lower Upper
 63 100
 112 0.0 2.9 4.3#



#59
 1,2,3-Trichloropropane
 Concen: 1.184 ug/L
 RT: 11.48 min Scan# 3231
 Delta R.T. 0.99 min
 Lab File: VU032445.D
 Acq: 04 Jun 2019 22:27

Tgt Ion: 75 Resp: 5894
 Ion Ratio Lower Upper
 75 100
 77 35.3 25.3 37.9

