

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060419\
 Data File : VU032445.D
 Acq On : 04 Jun 2019 22:27
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Jun 05 00:28:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 04 22:54:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	108296	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	99342	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	50529	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	30172	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.67	69	25374	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.40%
11) 1,1-Dichloroethene-d2	2.27	63	45569	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	4.18	46	79137	40.27	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.54%
24) Chloroform-d	4.64	84	52967	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
26) 1,2-Dichloroethane-d4	5.30	65	29723	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
32) Benzene-d6	5.33	84	111287	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
36) 1,2-Dichloropropane-d6	6.32	67	34656	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.56	98	102331	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	7.85	79	12661	3.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.80%
46) 2-Hexanone-d5	8.31	63	66685	40.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	28435	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	40844	4.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.80%

Target Compounds					Ovalue
34) Trichloroethene	6.18	95	1267	0.150 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060419\
Data File : VU032445.D
Acq On : 04 Jun 2019 22:27
Operator : JC/SP
Sample : VIBLK40
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jun 05 00:28:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 04 22:54:09 2019
Response via : Initial Calibration

