

Data File : VU038540.D
Acq On : 04 Jun 2020 10:43
Operator : SY/MD
Sample : VU0604WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK02

Quant Time: Jun 06 05:02:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 05 01:11:12 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	309515	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	312110	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	149826	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	107275	5.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.20%
7) Chloroethane-d5	1.93	69	85747	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.40%
11) 1,1-Dichloroethene-d2	2.59	63	163138	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.80%
20) 2-Butanone-d5	4.69	46	316899	52.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.16%
24) Chloroform-d	5.10	84	196836	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
26) 1,2-Dichloroethane-d4	5.74	65	115980	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.60%
32) Benzene-d6	5.76	84	409313	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.72	67	120904	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.92	98	348836	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	8.20	79	48973	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.65	63	267895	52.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.30%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	100422	4.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	132431	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU038540.D
Acq On : 04 Jun 2020 10:43
Operator : SY/MD
Sample : VU0604WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK02

Quant Time: Jun 06 05:02:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 05 01:11:12 2020
Response via : Initial Calibration

