

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060420\
 Data File : VU038546.D
 Acq On : 04 Jun 2020 14:10
 Operator : SY/MD
 Sample : L2856-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 05 02:08:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 05 01:11:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	372977	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	360772	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	170985	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	103736	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.93	69	80932	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	2.59	63	159835	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.69	46	309621	42.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.44%
24) Chloroform-d	5.10	84	189505	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.74	65	109555	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
32) Benzene-d6	5.76	84	400172	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.72	67	117574	4.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.00%
41) Toluene-d8	7.92	98	340785	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	8.20	79	48478	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.66	63	256410	43.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.18%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	94994	4.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	130192	4.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.20%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	3497	0.135 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060420\
Data File : VU038546.D
Acq On : 04 Jun 2020 14:10
Operator : SY/MD
Sample : L2856-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 05 02:08:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 05 01:11:12 2020
Response via : Initial Calibration

